

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION PROCEDURES – INTOXILYZER 9000

1. Prepare at least two simulators for use and allow them to warm up for at least thirty minutes prior to the first analysis. When changing solutions, allow the new solution to warm up for at least ten minutes after the heater light turns off for the first time. Ensure that each simulator maintains an air leak resistant seal and an operational temperature of 34 C (+/- 0.2 C).
2. Only distilled or deionized water will be used for the Alcohol Free Test and the Interferent Detect Test. Class A glassware must be used when measuring solutions.
3. Only approved and non-expired alcohol reference solution and non-expired dry gas standard from an approved source must be used during the applicable portions of the inspection.
4. Access the Level 3 Menu. Select Maintenance. Select Inspection Test. Enter Department Inspector last name, first name and middle initial at INSPECTOR prompts.
5. All results are reported to three decimal places in g/210L. The result must be 0.000 for each air blank. The instrument will abort the inspection process if the air blank result is not 0.000. Except for DIAGNOSTIC CHECK, if a check or test is out of compliance the instrument will prompt the Department Inspector to REPEAT (Y/N) the check or test. Each check or test may only be repeated once. If a check or test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 41a Department Inspection Report – Intoxilyzer 9000.
 - Verify DATE. Adjust if necessary. Verify TIME. Adjust if necessary.
 - Select MINIMUM SAMPLE VOLUME CHECK. When PROVIDE SAMPLE NOW is displayed, provide a breath sample volume of less than 1.1L as shown on the display. The instrument must again display PROVIDE SAMPLE NOW. Provide a breath sample volume of greater than 1.1L as shown on the display. The result must be 0.000. Enter Y/N at the OK prompt.
 - DIAGNOSTIC CHECK (Pre-Inspection). Select Diagnostic Pre-Inspection Diagnostic Check. The result must be OK for each diagnostic check. If any diagnostic check result is not OK, the instrument will abort the inspection process.
 - NUMBER OF SIMULATORS USED. Enter the number of simulators used during the inspection.
 - Select ALCOHOL FREE SUBJECT/MOUTH ALCOHOL TEST. When PROVIDE SAMPLE NOW is displayed, introduce an alcohol-free breath sample into the instrument. The result must be 0.000. Rinse mouth with mouth alcohol solution. When PROVIDE SAMPLE NOW is again displayed introduce a breath sample into the instrument. The result must be SLOPE NOT MET.
 - Select ALCOHOL FREE TEST. Attach a simulator containing 500 mL distilled or deionized water to the instrument. Conduct ten (10) analyses. The result must be 0.000 for each analysis.
 - Select INTERFERENT DETECT TEST. Attach a simulator containing 3 mL of acetone stock solution and 500 mL distilled or deionized water to the instrument. Conduct ten (10) analyses. The result must be INTERFERENT DETECT for each analysis.
 - Select 0.02 g/210L TEST. Attach a simulator containing 0.02 g/210L alcohol reference solution to the instrument. Enter the lot number and expiration date of the alcohol reference solution used. Conduct ten (10) analyses. The result of each analysis must be within the acceptable range.
 - Select 0.08 g/210L TEST. Attach a simulator containing 0.08 g/210L alcohol reference solution to the instrument. Enter the lot number and expiration date of the alcohol reference solution used. Conduct ten (10) analyses. The result of each analysis must be within the acceptable range.
 - Select 0.30 g/210L TEST. Attach a simulator containing 0.30 g/210L alcohol reference solution to the instrument. Enter the lot number and expiration date of the alcohol reference solution used. Conduct ten (10) analyses. The result of each analysis must be within the acceptable range.
 - Select 0.15g/210L DRY GAS STANDARD TEST. Attach a simulator containing 0.15 g/210L dry gas standard to the instrument. Enter the lot number and expiration date of the dry gas standard used. Conduct ten (10) analyses. The result of each analysis must be within the acceptable range.
 - Select 0.08 g/210L DRY GAS STANDARD TEST. Attach a cylinder containing 0.08 g/210L dry gas standard to the instrument. Enter the lot number and expiration date of the dry gas standard used. Conduct ten (10) analyses. The result of each analysis must be within the acceptable range.
 - Select DIAGNOSTIC CHECK (Post-Inspection). The result must be OK for each diagnostic check. If any diagnostic check result is not OK, the instrument will abort the inspection process.
 - REVIEW/ EDIT REMARKS if needed.
 - IN COMPLIANCE. Enter Y/N to record whether the instrument complies or does not comply with the requirements of Chapter 11D-8, FAC. If the instrument does not comply with Chapter 11D-8, FAC, remove the instrument from service and notify the Agency.
6. The results of the Department Inspection must be recorded on FDLE/ATP Form 41a Department Inspection Report – Intoxilyzer 9000. The standard deviations, and the average standard deviation, for the 0.02, 0.08 and 0.15, 0.30 g/210L